

Epileptic Detection and Classification using Convolutional Neural Network with Dual Tree Complex Wavelet Features

^aSreelatha Panda, ^bSatyasis Mishra, and, ^cMihir Narayan Mohanty

^a Rajadhani Engineering College

^b Dept. of ECE, Cenurion University of Technology and Management

^c Mihir Narayan Mohanty, ITER, Siksha 'O' Anusandhan (Deemed to be University)

Article History: Received: 10 January 2021; Revised: 12 February 2021; Accepted: 27 March 2021; Published online: 20 April 2021

Abstract: Epilepsy is a kind of brain disease that can be diagnosed by observation of EEG signals. Mostly it occurs within the children. However, some of the cases are observed in adults. It is a challenging task for physicians to detect this disease at an early stage. Authors in this work have classified the Epileptic and normal EEG signal by adopting the deep learning approach. For efficient features, dual tree complex wavelet (DTCWT) is considered. The decomposed wavelet features are used as the input to the convolutional neural network (CNN) classifier. Around 97% classification accuracy is observed by using the proposed approach.

1. Introduction

Human brain continuously generate the electrical impulses in a specific order pattern and these impulses travel among the neurons, nerve cells, and throughout the entire body via neurotransmitters. Generally, seizure is a type of brain disorder happens due to the immediate alternation and temporary changes of electrical activity in brain. Epileptic is a brain disorder that happens due to repeated seizures. Machine learning has a major role in biomedical data analysis and classification. Research on machine learning based biomedical data analysis is getting popular since last few decades. Deep learning is a most popular machine learning approach and it has been used in many applications due to its auto feature extraction nature. But sometimes it may omit the useful features from the data and it impact on classification accuracy. For this, a comparison of CNN for classifying the raw and decomposed EEG is shown in the proposed work. Numerous works were done related to epileptic detection and are discussed below.

In (Deshmukh P,2017) the framework for detection of epileptic seizures has been proposed. Discrete wavelet transform (DWT) was considered for analysis of EEG signals and feature extraction. They have considered 14 different combinations to classify the signal with two-class problem. Also the DWT co-efficient were used for eye opening and epileptic cases.5 Datasets have been used for the statical feature extraction from the DWT co-efficient. These features were fed to the KNN classifier and found the accuracy level. In (Geng D,2016)the accuracy of epileptic seizure detection and reduced computational cost has been focused. To improve the detection of accuracy and extract features different Datasets have been used by using 54 DWT Mother Wavelet. These features are reduced by using GA algorithm to select best features. Then it is fed to ANN classifier to perform better in terms of evaluation matrix. In (Mohapatra I,2019) the analysis framework for EEG signal is proposed. DWT was considered in which signals were decomposed into frequency sub-bands. Here statical features were extracted from sub-bands to represent the Wavelet co-efficient. The dimension of features wasreduced by using PCA, ICA and LDA.SVM has been used as a classifier with three features PCA, ICA& LDA. Authors in (Mohapatra S.K,2021), have considered wavelet neural network (WNN) for the purpose of detecting seizures from long duration EEG signal. Their proposed approach is a hybrid approach that combines conventional neural network with wavelet transform.

By using DWT relative amplitude and relative fluctuation index were extracted, and all features were extracted. At last, the feature vectors were fed into the trained classifier and found the accuracy level. In (Mardini W,2020) an effective method for Epileptic seizure was carried out. DWT was used for decomposition of sub-bands. A set of two Datasets were used for normal and seizure patients. Then features can be calculated from each sub band.

2. Proposed Approach

The proposed epileptic classification approach is a three stage model as shown in Figure 1. Physionet is an open source biomedical database from which the EEG data is taken for classification purpose. Relevant features are extracted by applying dual tree complex wavelet transform (DTCWT) in the original EEG. For classification, CNN is used that classifies the features extracted by DTCWT.




Figure 1. Proposed classification approach

2.1 Feature Extraction

For non-stationary signal analysis, wavelet transform plays an important role. It decomposes the input signal into various levels as presented in Figure 2. Generally, it decomposes the signal into detailed and approximate coefficients that corresponds to the low and high pass filtering. Wavelet transform is type of linear transform that filters the original signal to a multi-resolution representation by applying scaled and shifted approach of the mother wavelet. DTCWT is a class of wavelet transform where analytic filters are used for performing the wavelet decomposition. The real and imaginary parts are separated by DTCWT into two different trees. Here the signal is decomposed in terms of complex shifted and dilated mother wavelet $\psi(x)$, and scaling function $\phi(x)$. For a one dimensional input, the real and imaginary parts of the wavelet and scaling function can be derived by

$$\psi_r(t) = \sqrt{2} \sum_n H_a(n) \phi_r(2t - n)$$

$$\psi_i(t) = \sqrt{2} \sum_n H_b(n) \phi_i(2t - n)$$

$$\psi_r(t) = \sqrt{2} \sum_n L_a(n) \phi_r(2t - n)$$

$$\psi_i(t) = \sqrt{2} \sum_n L_b(n) \phi_i(2t - n)$$

where L and H represents the lowpass and highpass filters. The wavelet functions ψ_r and ψ_i produce the complex wavelet function ψ_c , which is given by $\psi_r + j\psi_i$ [6].




Figure 2. Structure of DTCWT decomposition process.

2.2 Neural Network (CNN)

CNN is a class of deep neural network architecture and has been widely used in different types of data analysis and classification. The data may be image, signal, attributes. Generally, a CNN is a neural network type classifier composed with number of convolutional layers. Each convolutional layer is followed by a pooling layer (sub-sampling). Again the pooling layer is followed by one or more fully connected layers. The structure of a CNN is presented in Fig. 3. There are also some certain numbers of neurons presented in both the convolutional and pooling layers. A weight sharing approach exists in between the neurons of both the layers to deal with the overfitting problem.




Figure 3. Structure of the CNN

The convolution process for the input data can be presented as follows:

$$\lambda_j^f = \omega^f n_j + \beta^f$$

$$\alpha_j^f = \frac{1}{1 + e^{-\lambda_j^f}}$$

where λ_j^f corresponds to the convolutional output result which is associated with j^{th} input data and feature map f . ω^f is the corresponding weight and β^f is the bias for the feature map f . α_j^f is the output for a non-linear activation function which is either sigmoid or tanh. n_j denotes the j th training data.

For reducing the dimensionality of the input data and increasing the learning process, maxpooling operation is applied to the hidden layer that is called sub sampling layer. The operation is performed by finding the maximum average of convolved feature among the adjacent neurons presented in preceding convolutional layer. The convolutional and sub sampling layers follow one or more fully connected layers are used whose neurons are connected to all the neurons from the preceding layer. Most of the CNN parameters are interested by the fully connected layer parameters. There are ten layers in the proposed one dimensional CNN classifier

including four convolutional layers, four max-pooling layer, and two fully connected layers. The detail summarization of the network structure is presented in Table II. After the convolutional process max pooling is applied for reducing the size of the feature set. Sigmoid activation function is applied in each convolutional layer (1, 3, 5, and 7). The fully connected neural network consists of 256, 128, and 64 neurons. The tanh activation function is used in the network for getting the separate output for each class. The network is trained by using the backpropagation algorithm. The weights and bias values are updated by using adaptive moment estimation (Adam) optimization algorithm This algorithm is a popular optimization algorithm is deep learning because it achieves good results fast(Sharmila A,2016). The complete structure of the proposed CNN classification model is presented in Fig. 4.




Figure 4. Proposed CNN structure.

TABLE II Layer summary of the proposed CNN

Layer Type	L ayer
Convolutional + Relu	0- 1
Max pooling	1- 2
Convolutional + Relu	2- 3
Max pooling	3- 4
Convolutional + Relu	4- 5
Max pooling	5- 6
Convolutional + Relu	6- 7
Fully connected	7- 8

Fully connected	8-9
Fully connected	9-10

3. Results and Discussion

The EEG signal is collected from Physionet open source repository. Normal and epileptic signals are shown in Figure 2 and 3. After collecting the data it is decomposed by wavelet transform. Decomposed signal are shown in Figure 4.




Figure 5: Normal EEG signal




Figure 6: Epileptic EEG signal




Figure 7: Decomposed EEG signal

The proposed CNN model is designed by using the Tensor flow framework in Python. The simulation process is held in an Intel Core i5 system with 8 GB RAM. The original data set is divided into training and testing set. About 90% (7675 samples) EEG data is considered for training and rest 10% data (853 samples) is taken for testing purpose of the network. CNN is trained with the training data by the backpropagation algorithm. The batch size for training is 275 and categorical cross-entropy is used as the loss function. Classification model is trained using backpropagation algorithm. The classification performance of the CNN classifier with DTCWT feature is presented in Fig. 14. The accuracy is quite better as compared to CNN without DTCWT.




Figure 8. Classification accuracy archived from CNN

The performance is also measured with different parameters as accuracy, specificity, and sensitivity. Sensitivity is the number of correct positive predictions divided by the total number of positives. Specificity is another performance measuring parameter and can be calculated as the number of correct negative predictions divided by the total number of negatives. Performance of the proposed method is presented in Table III and IV.

TABLE IV CNN+DTCWT classification result for each class.

Class	Sensitivity (%)	Specificity (%)	Accuracy (%)
Normal	99.40	98.27	99.29
Epileptic	96.26	99.87	99.76
Average	97.83	98.57	99.52

6. Conclusion

Automatic disease classification is one of the most important research in bioinformatic domain. A CNN classifier with DTCWT features is proposed to classify epileptic for early diagnosis. It is performing efficiently for classifying the EEG signals. The CNN with DTCWT features is performing better as compared to simple CNN model. A SVM classifier with gaussian kernel is designed for the validation of the result. Further the obtained result is also compared with some earlier works and it can be confirmed that the proposed method is performing effectively for EEG classification. Further the classification accuracy can be improved by optimising the network structure and that has kept as the future work.

References

1. Aradhyula, T.V., Bian, D., Reddy, A.B., Jeng, Y.R., Chavali, M., Sadiku, E.R. and Malkapuram, R., 2020. Compounding and the mechanical properties of catla fish scales reinforced-polypropylene composite—from biowaste to biomaterial. *Advanced Composite Materials*, 29(2), pp.115-128.
2. Arunkarthikeyan K., Balamurugan K. & Rao P.M.V (2020) Studies on cryogenically treated WC-Co insert at different soaking conditions, *Materials and Manufacturing Processes*, 35:5, 545-555, DOI: [10.1080/10426914.2020.1726945](https://doi.org/10.1080/10426914.2020.1726945)
3. Babu, U.V., Mani, M.N., Krishna, M.R. and Tejaswini, M., 2018. Data Preprocessing for Modelling the adulteration detection in Gasoline with BIS. *Materials Today: Proceedings*, 5(2), pp.4637-4645.
4. Deshmukh P, Ingle R, Kehri V, Awale, R. N. (2017) Epileptic seizure detection using discrete wavelet transform based support vector machine. In 2017 International Conference on Communication and Signal Processing (ICCSP) (pp. 1933-1937). IEEE.
5. Ezhilarasi, T.P., Kumar, N.S., Latchoumi, T.P. and Balayesu, N., 2021. A Secure Data Sharing Using IDSS CP-ABE in Cloud Storage. In *Advances in Industrial Automation and Smart Manufacturing* (pp. 1073-1085). Springer, Singapore.
6. Garikipati, P. and Balamurugan, K., 2021. Abrasive Water Jet Machining Studies on AlSi 7+ 63% SiC Hybrid Composite. In *Advances in Industrial Automation and Smart Manufacturing* (pp. 743-751). Springer, Singapore.
7. Geng D, Zhou W, Zhang Y, Geng S. (2016) Epileptic seizure detection based on improved wavelet neural networks in long-term intracranial EEG. *Biocybernetics and Biomedical Engineering*;36(2), 375-384.
8. Latchoumi, T.P., Dayanika, J. and Archana, G., 2021. A Comparative Study of Machine Learning Algorithms using Quick-Witted Diabetic Prevention. *Annals of the Romanian Society for Cell Biology*, pp.4249-4259.
9. Latchoumi, T.P., Vasanth, A.V., Bhavya, B., Viswanadapalli, A. and Jayanthiladevi, A., 2020, July. QoS parameters for Comparison and Performance Evaluation of Reactive protocols. In 2020 International Conference on Computational Intelligence for Smart Power System and Sustainable Energy (CISPSSE) (pp. 1-4). IEEE.
10. Mohapatra I, Pattnaik P, Mohanty, M.N. (2019) Cardiac Failure Detection Using Neural Network Model with Dual-Tree Complex Wavelet Transform. In *International Conference on Intelligent Computing and Applications* (pp. 87-95). Springer, Singapore.
11. Mohapatra S.K., Mohanty M.N. (2021) Convolutional Neural Network Based Arrhythmia Classification with Selective Features from Empirical Mode Decomposition. In *Proceedings of the Second International Conference on Information Management and Machine Intelligence* (pp. 375-383). Springer, Singapore.
12. Mardini W, Yassein M.M.B., Al-Rawashdeh R, Aljawarneh S, Khamayseh Y, Meqdadi O. (2020) Enhanced Detection of Epileptic Seizure Using EEG Signals in Combination with Machine Learning Classifiers. *IEEE Access*: 8:24046-24055.

13. Subasi A, Gursoy M.I. (2010). EEG signal classification using PCA, ICA, LDA and support vector machines. *Expert systems with applications*;37(12), 8659-8666.
14. Sharmila A, Geethanjali P. (2016) DWT based detection of epileptic seizure from EEG signals using naive Bayes and k-NN classifiers. *Ieee Access*; 4:7716-7727.
15. Srivatsav, V., Ravishankar, C., Ramakarishna, M., Jyothi, Y. and Bhanuparakash, T.N., 2018, August. Mechanical and thermal properties of chicken feather reinforced epoxy composite. In *AIP Conference Proceedings* (Vol. 1992, No. 1, p. 040034). AIP Publishing LLC.
16. Yarlagaaddaa, J. and Malkapuram, R., 2020. Influence of carbon nanotubes/graphene nanoparticles on the mechanical and morphological properties of glass woven fabric epoxy composites. *INCAS Bulletin*, 12(4), pp.209-218.
17. Yookesh, T.L., Boobalan, E.D. and Latchoumi, T.P., 2020, March. Variational Iteration Method to Deal with Time Delay Differential Equations under Uncertainty Conditions. In 2020 International Conference on Emerging Smart Computing and Informatics (ESCI) (pp. 252-256). IEEE.